

An International Collaboration to Harmonize the Quantitative Plasma Epstein-Barr Virus DNA Assay for Future Biomarker-Guided Trials in Nasopharyngeal Carcinoma

Quynh-Thu Le¹, Qiang Zhang², Hongbin Cao¹, Ann-Joy Cheng³, Benjamin A. Pinsky¹, Ruey-Long Hong⁴, Joseph T. Chang³, Chun-Wei Wang⁴, Kuo-Chien Tsao³, YM Dennis Lo⁶, Nancy Lee⁵, K. Kian Ang⁷, Anthony T.C. Chan⁶, and K.C. Allen Chan⁶

Abstract

Purpose: Persistently elevated posttreatment plasma EBV DNA is a robust predictor of relapse in nasopharyngeal carcinoma (NPC). However, assay standardization is necessary for use in biomarker-driven trials. We conducted a study to harmonize the method between four centers with expertise in EBV DNA quantitation.

Experimental Design: Plasma samples of 40 patients with NPC were distributed to four centers. DNA was extracted and EBV DNA copy number was determined by real-time quantitative PCR (*Bam*HI-W primer/probe). Centers used the same protocol but generated their own calibrators. A harmonization study was then conducted using the same calibrators and PCR master mix and validated with ten pooled samples.

Results: The initial intraclass correlations (ICC) for the first 40 samples between each center and the index center were 0.62 [95% confidence interval (CI): 0.39–0.78], 0.70 (0.50–0.83), and 0.59 (0.35–0.76). The largest variability was the use of different PCR master mixes and calibrators. Standardization improved ICC to 0.83 (0.5–0.95), 0.95 (0.83–0.99) and 0.96 (0.86–0.99), respectively, for ten archival frozen samples. For fresh plasma with spiked-in EBV DNA, correlations were more than 0.99 between the centers. At 5 EBV DNA copies per reaction or above, the coefficient of variance (CV) was less than 10% for the cycle threshold (C_T) among all centers, suggesting this concentration can be reliably used as a cutoff for defining the presence of detectable EBV DNA.

Conclusions: Quantitative PCR assays, even when conducted in experienced clinical labs, can yield large variability in plasma EBV DNA copy numbers without harmonization. The use of common calibrators and PCR master mix can help to reduce variability. *Clin Cancer Res*; 19(8); 2208–15. ©2013 AACR.

Introduction

Treatment with concurrent cisplatin (CDDP)-based chemoradiotherapy (CCRT) followed by 3 cycles of adjuvant CDDP and 5-fluorouracil (5-FU) is the current standard of care in the United States for locally advanced nasopharyngeal carcinoma (NPC; ref. 1). Recent advances in radiation delivery and the use of concurrent chemotherapy have

substantially increased the rate of local control, now ranging between 91% and 96% (2–6). However, the development of distant metastases remains problematic (~30% at 5 years) and ultimately results in death (4–6). Compliance to adjuvant chemotherapy after CCRT is problematic as only half of patients completed all adjuvant chemotherapy due to severe toxicity (1, 7). The results of a recent randomized trial from China, which failed to show a survival advantage for adjuvant chemotherapy (8) questioned benefit of adjuvant CDDP/FU. This study was criticized for not using a noninferiority design, and hence was potentially underpowered. In contrast, another study suggested that adjuvant 5-FU chemotherapy decreases distant metastasis in NPC (9). A plausible contributor to these conflicting findings is the inability to properly classify patients with different risk profile for enrollment in trials. The development of a biomarker to reliably identify the subset of patients at high risk for metastasis may help to identify patients that would benefit from adjuvant chemotherapy while sparing low-risk patients from unnecessary toxic treatment.

Blood is in direct contact with all organs and is an attractive sample type for noninvasive cancer surveillance.

Authors' Affiliations: ¹Stanford University, Stanford, California; ²RTOG Statistical Center, Philadelphia, Pennsylvania; ³Chang-Gung University, Taoyuan; ⁴National Taiwan University Hospital, Taipei, Taiwan, Republic of China; ⁵Memorial Sloan Kettering Cancer Center, New York, New York; ⁶State Key Laboratory in Oncology in South China, The Chinese University of Hong Kong, Hong Kong, S.A.R. China; and ⁷M.D. Anderson Cancer Center, Houston, Texas

Note: Supplementary data for this article are available at Clinical Cancer Research Online (<http://clincancerres.aacrjournals.org>).

Corresponding Author: Quynh-Thu Le, Department of Radiation Oncology, Stanford University, 875 Blake Wilbur Dr., MC 5847, Stanford, CA 94305. Phone: 650-498-5032; Fax: 650-735-8231; E-mail: Qle@stanford.edu

doi: 10.1158/1078-0432.CCR-12-3702

©2013 American Association for Cancer Research.

Translational Relevance

This is a study to harmonize the measurement of the plasma biomarker EBV DNA in four accredited international labs in order to launch a biomarker-driven international phase III trial in nasopharyngeal carcinoma (NPC). Although EBV DNA is a well-accepted robust prognostic marker in NPC and has been offered in several clinical laboratories as a means to track tumor burden and posttreatment surveillance, little is known about the interlaboratory variability of this quantitative assay. In this study, we showed that the interlaboratory variability is quite large for the same assay using identical procedures and primer/probe set without harmonization. We showed that harmonization, which involves standardization of buffers and calibrators, is feasible and significantly reduces such variability. Through this harmonization process, we established a standardized assay that can be used internationally for the measurement of this biomarker for future prospective studies and developed a process for credentialing new laboratories.

As the first evidence showing that tumor-associated DNA can be detected in the blood (10), several studies have evaluated tumor DNA as biomarkers for cancer surveillance (11–16). The presence of viral DNA in viral-related tumors offers a distinct marker for detection in the blood. Epstein Barr Virus (EBV) DNA is often found in the plasma of patients with NPC and has been shown to be a reliable marker for prognostication in this disease (14, 15, 17). Specifically, pretreatment plasma EBV DNA correlates with cancer stage (17, 18) and clinical outcome (19) in endemic NPC. Posttreatment (radiation or CCRT) plasma EBV DNA, defined as undetectable, has an even better correlation with prognosis and has been used to monitor recurrence after therapy (17, 18, 20, 21). Although undetectable posttreatment plasma EBV DNA was associated with an excellent progression-free survival (PFS: 80%–90%), persistently detectable level was associated with an extremely poor PFS (10%–15%) and may be a marker of subclinical residual disease (21). These observations have been consistently reproduced in large patient cohorts treated in different countries from both endemic and non-endemic areas (18, 20, 21). Published data to date indicate that the most robust and reliable biomarker for NPC prognostication is the posttreatment EBV DNA level.

Given the robustness of posttreatment plasma EBV DNA in prognostication, we propose to incorporate this biomarker in the next RTOG (Radiation Therapy Oncology Group) phase III NPC trial. We postulate that patients with undetectable EBV DNA after CCRT have a low risk for distant relapse and will be randomized to receive adjuvant cisplatin/5FU (current standard) versus observation. In contrast, those who continue to have detectable EBV DNA levels are at a "high risk" for distant relapse and will be randomized to cisplatin/5-FU (current standard) versus a

more intensified regimen. Because this will be an international trial enrolling patients from different countries, the logistical difficulty and high cost associated with shipping plasma samples across the continents, as well as the need for rapid result generation for randomization, it is important to first prospectively validate the assay for EBV DNA and to harmonize it across the different clinical laboratories in participating countries. All four participating laboratories have had extensive experience in measuring plasma EBV DNA in NPC patients and are nationally accredited. Here, we report the results of a prospective effort to harmonize this assay across these four international laboratories in preparation for the upcoming trial.

Although different primer/probe sets have been used to measure circulating EBV DNA in the plasma or serum, the most commonly used primer/probe set is the one targeting the *Bam*HI-W region of the EBV genome (14). This fragment occurs 8 to 11 times in the EBV genome, allowing more sensitive detection when compared to a single copy EBV genes, such as EBNA1, LMP2, or POL1 (18, 22). Because the purpose of the phase III trial is to identify patients with detectable EBV DNA at diagnosis for entry into the trial, and to distinguish patients with detectable levels from those with undetectable levels after chemoradiation for risk stratification and treatment assignment, it is critical that the most sensitive assay be used. More importantly, the largest and most robust published studies that established the prognostic significance of posttreatment circulating EBV DNA in NPC employed the *Bam*HI-W primer/probe set (20, 21, 23). Therefore, we decided to use this assay for EBV DNA measurement for the upcoming trial and focus our efforts to harmonize the assay in the participating laboratories.

Materials and Methods

Participating laboratories

Four laboratories that were selected for this harmonization include: (i) the Stanford Clinical Virology Laboratory [STF, certified under the Clinical Lab Improvement Amendment (CLIA)], which will serve as the central laboratory for the United States sites, (ii) the Chemical Pathology laboratory at The Chinese University of Hong Kong, Prince of Wales Hospital, Hong Kong [accredited by the National Association of Testing Authorities (NATA) and the Royal College of Pathologist of Australia], which will serve as the central laboratory for Hong Kong sites, (iii) the National Taiwan University Hospital Clinical Laboratory [NTU, Taiwan Accreditation Foundation (TAF) and participated in the College of American Pathologist (CAP) proficiency program] and (iv) the Chang-Gung Memorial Hospital-Linkou Clinical laboratory (CG, accredited by both TAF and CAP), which will serve as the central laboratories for Taiwanese sites. These laboratories were selected because of their commitment to the planned international phase III RTOG trial mentioned above, their experience in measuring plasma EBV DNA, their accreditation status, and their ability to offer the test as a clinical assay.

Patient samples

For the preharmonization study, anonymized plasma samples from 40 newly diagnosed NPC patients were collected from CG with patient consent and distributed to the 4 laboratories. Patients with all different stages (22 stage I and II and 18 stage III and IV) were included. For the harmonization process, 23 plasma samples of newly diagnosed NPC patients were collected from STF under an institutional review board approved study (NCT00186433), pooled and distributed to the 4 laboratories. For the postharmonization validation study, plasma samples from anonymized 40 patients with NPC were combined to create 10 pooled samples, aliquoted, and distributed.

DNA extraction

A total of 3.5 mL of blood was collected into EDTA-coated tubes, centrifuged for 10 minutes, plasma recovered, and frozen at -80°C until shipped. DNA was isolated manually from 400 μL of plasma using the QIAamp Blood Mini Kit (QIAGEN Inc), eluted with 50 μL of elution buffer or water. For the harmonization, 2 different operators from the same laboratory conducted DNA extraction and quantitative qPCR to determine interoperator variability.

Calibrators for standard curves

Calibrators were prepared by using DNA extracted from an EBV-positive cell line Namalwa, which is a diploid cell line containing 2 integrated viral genomes per cell as previously described (14). This DNA was also used for the spiking experiments.

Real-time quantitative PCR assay

DNA samples were quantified for EBV DNA using a real-time quantitative (qPCR) system targeting the *Bam*HI-W fragment region of the EBV genome as described by Lo and colleagues (14). Each run included patient samples, calibrators for constructing a standard curve, and appropriate positive, negative, and no template controls. All PCRs were performed in triplicate. The following real-time PCR detector systems were used: STF—Rotor-gene Q (Qiagen) and ABI7900; HK—ABI7300; NTU—ABI7900HT; CG—ABI7900 (Applied Biosystems). The plasma concentration of EBV DNA (copies/mL) was calculated as previously described (14).

Statistical analysis

All assay results are summarized using mean, SDs, and log-transformed if the data deviates from normality assumptions. Inter-rater reliability is estimated by intraclass correlation coefficient (ICC), a method to assess reproducibility of assay measures among different laboratories. On the basis of Shoukri and colleagues (24), with one-sided 5% type I error rate and 80% power, we need 39 patients to detect a difference of 0.2 assuming ICC of 0.6 under the null hypothesis between 2 laboratories. The within and between subject variations and ICCs are estimated using a general linear model with measurement error (25). The interlaboratory variability in qPCR measurement and DNA extraction was also summarized using the general linear model. Spear-

man rank correlation coefficient was used to assess results for the "spike-in" experiment among the 4 laboratories. All analyses were conducted using SAS 9.2.

Results

Because of the small volume of plasma collected per patient, we used 400 μL of plasma for DNA extraction

Table 1. Quantitation of EBV DNA copies/mL of plasma in 40 patients with NPC by 4 different clinical laboratories before harmonization

Patient number	NTU copies/mL	CG copies/mL	HK copies/mL	STF copies/mL
1	0	0	0	0
2	0	0	0	0
3	0	0	57	73
4	0	25	46	23
5	0	25	47	45
6	0	50	80	106
7	0	50	161	0
8	0	50	254	8
9	0	50	283	17
10	0	50	349	196
11	0	100	0	132
12	0	100	679	211
13	0	125	742	NA
14	0	150	279	0
15	0	175	300	151
16	0	325	199	321
17	0	725	1200	466
18	21	175	5600	1800
19	24	250	661	374
20	38	125	1579	403
21	38	75	102	149
22	54	75	559	475
23	112	700	4308	1333
24	139	325	1184	NA
25	140	925	4141	78
26	156	725	2386	139
27	173	1250	3410	NA
28	341	2225	7367	4003
29	369	1475	3600	NA
30	720	2700	10431	4906
31	946	2575	19410	552
32	1637	3275	20402	3697
33	2098	17850	20286	6363
34	3003	11450	41087	15392
35	3227	14900	78465	8136
36	8304	80000	359045	71956
37	8728	17575	23093	14029
38	18713	101825	145163	22786
39	30014	192325	375438	167832
40	82378	105900	4013627	22931

Abbreviation: NA, not assessable.

Table 2. ICC for each site when compared to the index site before and after the harmonization process

Site	Pre-harmonization	Post harmonization
	ICC (95% confidence interval) N = 40	ICC (95% confidence interval) N = 10
NTU vs. STF	0.62 (0.39–0.78)	0.83 (0.50–0.95)
CG vs. STF	0.70 (0.50–0.83)	0.72 (0.26–0.91) ^a
		0.95 (0.83–0.99) ^b
HK vs. STF	0.59 (0.35–0.76)	0.96 (0.86–0.99)

^aAggregated samples.^bNonaggregated samples.

instead of the 800 μ L that is normally used in the clinical protocol. The results for the initial run using 40 patient samples are shown in Table 1. There was a large interlaboratory variability in the absolute number of EBV DNA copies/mL. The detection rate for EBV DNA, defined here as ≥ 0 copy/mL was 58% for NTU, 80% for STF, and 93% for both CG and HK. Table 2 shows the ICC for each site compared with STF as the index site. All correlations had less than the desired 0.80 value. The observed variability between centers was not related to differences in qPCR instruments because in one lab (STF), we tested these 40 samples on the Rotor-Gene Q and ABI7900 and observed similar results for both instruments (data not shown).

To harmonize the assay, we identified nonstandardized factors that could be modified; these included calibrators (for standard curve), which were individually prepared in each laboratory, and the TaqMan master mix, which was purchased from 2 different vendors. Three laboratories used the premade Roche master mix (Roche Applied Sciences), whereas one laboratory prepared the master mix in-house using components from ABI (Applied Biosystems). As shown in Table 3, the use of different master mixes, when controlled for other aspects of the procedure, including the DNA extraction, operators, other reagents, calibrators and PCR instruments, resulted in a large difference in the measured EBV DNA copy number. Laboratories using the Roche master mix gave similar results; whereas the laboratory using PCR kits from ABI gave results 5- to 10-fold higher. Therefore, standardization was made to use the Roche master mix in subsequent studies.

Next, we assessed the variability in the qPCR step and DNA extraction by different operators within the same laboratory and by different calibrator sets. To test the interlaboratory variability in qPCR, we provided the 4 laboratories with the same amount of extracted EBV DNA from pooled NPC plasma samples at the concentration of approximately 4000 copies/mL. In one laboratory (HK), the results were available only for one calibrator set. As shown in the top half of Table 4, qPCR variability was much larger for the different calibrator sets than for the different operators. Therefore, we standardized the calibrators using those prepared by HK in all laboratories for the remaining harmonization process. To ensure that calibrator shipping did not result in degradation, we shipped a calibrator set from Hong Kong to the United States then back to Hong

Table 3. Ct values and calculated EBV copies for the calibrators and the plasma samples when two different PCR master mixes were used

Calibrator	Prepared Master Mix from ABI components				Roche Master Mix			
	Expected value	Observed read 1		Observed read 2		Observed read 1		Observed read 2
Copies/5 μ L	Copies/5 μ L	C _t	Copies/5 μ L	C _t	Copies/5 μ L	C _t	Copies/5 μ L	C _t
15000	60,456	23.77	59,454	23.79	13,153	23.28	13,001	23.30
6000	36,283	24.45	33,594	24.56	5,645	24.64	5,362	24.72
600	4,744	27.16	4,306	27.29	551	28.37	590	28.26
120	933	29.33	883	29.40	130	30.70	112	30.93
60	311	30.79	382	30.52	75	31.57	63	31.86
30	158	31.69	181	31.51	24	33.38	32	32.93
15	82	32.55	78	32.63	13	34.42	10	34.77
7.5	26	34.09	14	34.92	8	35.13	10	34.83
3.75	17	34.65	9	35.57	4	36.23	6	35.69
	Copies/mL	C _t			Copies/mL	C _t		
Plasma 1	41,831	28.55			4638	30.12		
Plasma 2	37,349	28.70			4432	30.19		
Plasma 3	43,812	28.49			4902	30.03		
Plasma 4	50,680	28.30			4849	30.05		

NOTE: Expected and observed values are shown. Note that the expected values for all 4 plasma samples were \sim 4,000 copies/mL.

Table 4. Two different sets of calibrators were used to measure EBV DNA from extracted DNA or plasma samples with expected values of 4000 copies/mL

Samples	Site	CG Calibrator		HK Calibrator	
		Operator 1	Operator 2	Operator 1	Operator 2
DNA (expected value: 4000 copies/mL)	STF	2775 (270)	2533 (118)	3358 (347)	3067 (146)
	NTU	3288 (21)	3089 (63)	4108 (27)	4300 (240)
	CG	2833 (298)	2708 (505)	4525 (499)	3675 (730)
	HK	NA	NA	4162 (157)	4034 (40)
Plasma (expected value: 4000 copies/mL)	STF Op1			2738 (197)	2538 (319)
	STF Op2			2654 (296)	2388 (406)
	NTU Op1			3462 (212)	3516 (227)
	NTU Op2			3653 (259)	4121 (411)
	CG			6030 (1626)	4160 ^a
	HK			4338 (141)	4671 (36)

NOTE: Operator 1 and 2 in the second row reflects different qPCR operators. Op1 and Op2 in the second column refer to different operators for DNA extraction. The mean concentrations are presented and the numbers in parenthesis represent the SD.

^aOnly one value of the triplicate was interpretable for this reading.

Kong and tested it against the same batch that was not shipped. Shipping did not result in a significant decrease in calibrator performance (Supplementary Table S1).

We also evaluated the DNA extraction variability between the 4 different laboratories, using the same DNA extraction kit. In 2 laboratories (STF and NTU), we also assessed DNA extraction variability by different operators. For this analysis, aliquots of a pooled plasma sample with EBV DNA concentration of approximately 4,000 copies/mL were distributed to the four laboratories. As shown in the lower half of Table 4, there was minimal interoperator variability. There was likewise good agreement between the 4 laboratories and the difference was within one PCR threshold cycle (data not shown).

We then validated the harmonization with all laboratories using the HK calibrators, Roche master mix and standardized procedures. Because fresh plasma samples were not available, we pooled 40 archival NPC plasma samples that had previously been frozen and thawed at least once to generate 10 pooled samples with different concentrations. The samples were shipped from Hong Kong to the United States and then to Taiwan. As shown in Table 2, the ICC improved from 0.62 to 0.83 for NTU versus STF; 0.70 to 0.72 for CG versus STF, and 0.59 to 0.96 for HK versus STF. Interestingly, there was significant protein aggregation noted in the plasma samples received by the CG site, presumably due to prolonged stay at room temperature related to delayed shipment delivery. Measurements of EBV DNA from the protein aggregate and the supernatant from the same plasma samples yielded significantly different results with the aggregate showing 2.3 to 2.5 times higher copy number than the supernatant. Therefore, the CG used the left over samples received by the other Taiwanese site, NTU, which did not have much aggregation. Results from this repeated assay showed an ICC of 0.95 between CG versus STF (Table 2). As all laboratories will be testing fresh plasma

samples without protein aggregation for the trial, another way to validate the harmonization is to "spike-in" known concentrations of EBV DNA into fresh, negative, non-NPC plasma samples. Table 5 shows the result of the "spike-in" experiment, which was highly consistent between the 4 laboratories. The correlations were >0.99 ($P < 0.0001$) between Stanford and the other 3 laboratories.

Because an important aspect of the planned phase III trial is to distinguish patients with detectable and undetectable post CCRT plasma EBV DNA for risk stratification, it is important to show that the harmonized assay is sensitive enough to measure a very low level of EBV DNA in the plasma. As there are 8 to 11 units of the *BamHI-W* fragment in each EBV genome, this assay is more sensitive than assays

Table 5. EBV DNA "spike in" experiment

Spike-in DNA Concentration	STF	NTU	CG	HK
None	0	0	0	0
600	720	336	335	519
6000	3445	4589	2830	6115
15000	13000	14493	11100	14087
Correlation		>0.99	>0.99	>0.99
Coefficient ^a				
<i>P</i>		0.0058	0.0030	0.0034

NOTE: EBV DNA at different concentrations was added to fresh plasma samples from an EBV DNA-negative, non-NPC patient to simulate fresh plasma from a NPC patient. Extraction and quantitation was performed in each laboratory using the harmonized assay.

^aCorrelation was calculated between each site and STF value, after log transformation, with 0 value set to 0 log value.

detecting targets with a single copy. To investigate the detection limit of this assay for all involved laboratories, we analyzed 10 to 20 replicates of diluted DNA from the Namalwa cell line at a concentration of 0.5, 1.25, 2.5, 5, 25, and 100 copies per reaction. Although the assay showed positive signals in several replicates at concentrations below 5 copies per reaction, the CV for the number of PCR threshold cycle (C_t) was greater than the 10% that is normally accepted for a clinical test. At 5 copies per reaction, the CV was consistently less than 10% for all 4 sites. However, even at these low CVs, the SD for C_t can be up to 1.1 cycles. If we use a fixed C_t threshold (mean or median value), up to approximately 50% of the samples having that concentration would be falsely excluded. Therefore, we decided to use the mean C_t value + 2 SDs at the concentration of 5 copies per reaction as a cutoff for defining a detectable level in the subsequent clinical trial. Theoretically, this would include 95% of the samples having an actual concentration of 5 copies/reaction, which translates to 60 copies/mL of patient plasma.

Discussion

To conduct a biomarker-driven study, it is crucial that the assay for the biomarker is performed in a central laboratory with CLIA or equivalent certification, which applies to all participating laboratories here. However, in situations where it is not feasible to use one central laboratory due to the size of the study and the logistical/cost issues of shipping fresh samples across continents, it is important that the assay be standardized across the participating laboratories. Although plasma EBV DNA is a well-known prognostic marker for NPC and is currently being offered as a clinical test at several institutions, little is known about the inter-laboratory variability of the assay.

Here, we showed that the intraclass variability between the different clinical laboratories with significant experience in this assay could be quite large without harmonization. Major contributors to the inter-laboratory variability were the PCR master mix and calibrators, more so than interoperator variability. Surprisingly, different PCR master mixes yielded more than 5-fold divergence in EBV DNA measurements, despite using the same calibrators, suggesting a difference in amplification efficiencies between plasma DNA and calibrators. Similarly, different calibrator sets, though prepared from the same cell line and protocol, resulted in larger variability than interoperator variability. Hence, harmonization using the same calibrators and PCR master mixes should result in less interlaboratory inconsistency. For the planned clinical trial, calibrators will be prepared in HK and shipped to all sites. Fresh calibrators will also be calibrated against old ones to maintain consistency over time. Shipping calibrators on dry ice did not result in degradation or affect calibrator performance.

The first WHO International Standard for EBV for Nucleic Acid Amplification Techniques NIBSC code: 09/260 became available during the course of this study. Given the

current efforts to harmonize this assay, we anticipate that additional harmonization using World Health Organization (WHO) material will further improve correlation and allow the results of future biomarker trials to be more generalizable. Therefore, we plan to use both the WHO standards and the Namalwa DNA as calibrators during the trial to prospectively compare the performance of both standards in a large NPC patient group.

Although harmonization resulted in ICC improvement for all sites, it was less marked for the Taiwanese sites compared to that between STF and HK. Our detailed analyses indicated that prolonged exposure to room temperature of previously frozen plasma samples resulted in marked protein aggregation, which surprisingly influenced the readings. Higher levels were noted from the protein aggregate than the supernatant from the same plasma samples. In contrast, samples without aggregation and the use of "spike-in" samples, which closely resembled fresh plasma, showed minimal variability between the laboratories, confirming the success of the harmonization.

Because of (i) the sensitive nature of qPCR, (ii) the amplification factor used to convert copies per reaction to copies/mL that can magnify interoperator and interlaboratory variability, and (iii) C_t being the direct assay readout, we believe that the C_t value is more useful than copy/mL in defining the minimum detection limit of this assay for risk stratification. Although two (STF and HK) laboratories were able to detect EBV DNA at a concentration as low as 0.5 copy/reaction (i.e., 6 copies/mL), the detectability of EBV DNA at this level was relatively unpredictable for the other 2 laboratories (CG and NTU). Therefore, if we simply use any detectable level of EBV DNA as the criteria for risk stratification, significant intercenter variation would be expected. On the other hand, if we use a fixed quantitative cutoff value, the variation in quantitative measurement may also result in significant inter-center differences. For example, if we measure 100 plasma aliquots, each with a putative EBV DNA concentration of 500 copies/mL, and use the median measured concentration as a cutoff, then 50 aliquots would have a measured concentration above and 50 would have a measured concentration below the cutoff. The samples with measured concentration below the cutoff value by random variation would be falsely rejected. To resolve this potential confounding issue, we propose to use a cutoff value derived from the mean C_t of a concentration that all 4 laboratories can consistently detect (≥ 5 copies/reaction) plus 2 SDs. Using this cutoff ensures that 95% of the samples having an actual concentration of 5 copies/reaction or 60 copies/mL would be correctly classified as detectable. To obtain the cutoff for each run, the laboratory will include 10 to 20 replicates of 5 copies per reaction to accurately determine the mean and SD for the C_t value at this concentration; the detectability point will be at 2 SDs above the mean C_t value. Plasma samples having a C_t below this cutoff (corresponding to a higher EBV DNA copy number) will be regarded as having a detectable level of EBV DNA.

In summary, we detected significant variability in plasma EBV DNA measurements between different clinical laboratories, which substantially improved with harmonization. This establishes a standardized assay that can be used internationally for the measurement of this biomarker for future prospective studies. It also provides a process for credentialing new laboratories, and ensures that the trial results will be applicable to the real world. The development of clinically actionable biomarkers is the key to personalized medicine and this harmonization is an important benchmark for all future biomarker-driven studies.

Disclosure of Potential Conflicts of Interest

Y.M.D. Lo is a consultant/advisory board member of Sequenom. No potential conflicts of interest were disclosed by the other authors.

Authors' Contributions

Conception and design: Q.-T. Le, Q. Zhang, H. Cao, N.Y. Lee, K.-K. Ang, A.T.C. Chan, K.C.A. Chan

Development of methodology: Q.-T. Le, H. Cao, K.-C. Tsao, Y.M.D. Lo, K.C.A. Chan

Acquisition of data (provided animals, acquired and managed patients, provided facilities, etc.): Q.-T. Le, A.-J. Cheng, B.A. Pinsky, R.-L. Hong, J.T.-C. Chang, J.T.-C. Chang, C.-W. Wang, K.-C. Tsao, N.Y. Lee, A.T.C. Chan, K.C.A. Chan

Analysis and interpretation of data (e.g., statistical analysis, biostatistics, computational analysis): Q.-T. Le, Q. Zhang, H. Cao, A.-J. Cheng, B.A. Pinsky, K.-C. Tsao, K.-K. Ang, K.C.A. Chan

Writing, review, and/or revision of the manuscript: Q.-T. Le, Q. Zhang, B.A. Pinsky, J.T.-C. Chang, C.-W. Wang, Y.M.D. Lo, N.Y. Lee, K.-K. Ang, A.T.C. Chan, K.C.A. Chan

Administrative, technical, or material support (i.e., reporting or organizing data, constructing databases): J.T.-C. Chang, Y.M.D. Lo, A.T.C. Chan

Study supervision: Q.-T. Le, K.-C. Tsao, Y.M.D. Lo, A.T.C. Chan

Grant Support

This study was supported by grants U10 CA21661 from the National Cancer Institute (NCI), the Li Ka Shing Foundation (to Y.M.D. Lo, A.T.C. Chan, and K.C.A. Chan).

The costs of publication of this article were defrayed in part by the payment of page charges. This article must therefore be hereby marked *advertisement* in accordance with 18 U.S.C. Section 1734 solely to indicate this fact.

Received November 30, 2012; revised February 6, 2013; accepted February 6, 2013; published OnlineFirst March 4, 2013.

References

- Al-Sarraf M, LeBlanc M, Giri PG, Fu KK, Cooper J, Vuong T, et al. Chemoradiotherapy versus radiotherapy in patients with advanced nasopharyngeal cancer: a phase III randomized intergroup study 0099. *J Clin Oncol* 1998;16:1310-7.
- Lee N, Xia P, Quivey JM, Sultanem K, Poon I, Akazawa C, et al. Intensity-modulated radiotherapy in the treatment of nasopharyngeal carcinoma: an update of the UCSF experience. [see comment]. *Int J Radiat Oncol Biol Phys* 2002;53:12-22.
- Lee N, Harris J, Garden AS, Straube W, Glisson B, Xia P, et al. Intensity-modulated radiation therapy with or without chemotherapy for nasopharyngeal carcinoma: radiation therapy oncology group phase II trial 0225. *J Clin Oncol* 2009;27:3684-90.
- Kam MK, Teo PM, Chau RM, Cheung KY, Choi PH, Kwan WH, et al. Treatment of nasopharyngeal carcinoma with intensity-modulated radiotherapy: the Hong Kong experience. *Int J Radiat Oncol Biol Phys* 2004;60:1440-50.
- Kwong DL, Pow EH, Sham JS, McMillan AS, Leung LH, Leung WK, et al. Intensity-modulated radiotherapy for early-stage nasopharyngeal carcinoma: a prospective study on disease control and preservation of salivary function. *Cancer* 2004;101:1584-93.
- Wolden SL, Chen WC, Pfister DG, Kraus DH, Berry SL, Zelefsky MJ. Intensity-modulated radiation therapy (IMRT) for nasopharynx cancer: update of the Memorial Sloan-Kettering experience. *Int J Radiat Oncol Biol Phys* 2006;64:57-62.
- Lee AW, Lau WH, Tung SY, Chua DT, Chappell R, Xu L, et al. Preliminary results of a randomized study on therapeutic gain by concurrent chemotherapy for regionally-advanced nasopharyngeal carcinoma: NPC-9901 Trial by the Hong Kong Nasopharyngeal Cancer Study Group. *J Clin Oncol* 2005;23:6966-75.
- Chen L, Hu CS, Chen XZ, Hu GQ, Cheng ZB, Sun Y, et al. Concurrent chemoradiotherapy plus adjuvant chemotherapy versus concurrent chemoradiotherapy alone in patients with locoregionally advanced nasopharyngeal carcinoma: a phase 3 multicentre randomised controlled trial. *Lancet Oncol* 2012;13:163-71.
- Lee AW, Tung SY, Ngan RK, Chappell R, Chua DT, Lu TX, et al. Factors contributing to the efficacy of concurrent-adjuvant chemotherapy for locoregionally advanced nasopharyngeal carcinoma: combined analyses of NPC-9901 and NPC-9902 Trials. *Eur J Cancer* 2011;47:656-66.
- Leon SA, Shapiro B, Sklaroff DM, Yaros MJ. Free DNA in the serum of cancer patients and the effect of therapy. *Cancer Res* 1977;37:646-50.
- Usadel H, Brabender J, Danenberg KD, Jeronimo C, Harden S, Engles J, et al. Quantitative adenomatous polyposis coli promoter methylation analysis in tumor tissue, serum, and plasma DNA of patients with lung cancer. *Cancer Res* 2002;62:371-5.
- Shao ZM, Wu J, Shen ZZ, Nguyen M. p53 mutation in plasma DNA and its prognostic value in breast cancer patients. *Clin Cancer Res* 2001;7:2222-7.
- Anker P, Mulcahy H, Stroun M. Circulating nucleic acids in plasma and serum as a noninvasive investigation for cancer: time for large-scale clinical studies? *Int J Cancer* 2003;103:149-52.
- Lo YM, Chan LY, Lo KW, Leung SF, Zhang J, Chan AT, et al. Quantitative analysis of cell-free Epstein-Barr virus DNA in plasma of patients with nasopharyngeal carcinoma. *Cancer Res* 1999;59:1188-91.
- Mutirangura A, Pornthanakasem W, Theamboonlers A, Sriuranpong V, Lertsanguansinchi P, Yenrudi S, et al. Epstein-Barr viral DNA in serum of patients with nasopharyngeal carcinoma. *Clin Cancer Res* 1998;4:665-9.
- Utting M, Werner W, Muller G, Schubert J, Junker K. A possible noninvasive method for the detection of bladder cancer in patients: microsatellite analysis of free DNA in urine and blood. *Ann NY Acad Sci* 2001;945:31-5.
- Lo YM, Chan LY, Chan AT, Leung SF, Lo KW, Zhang J, et al. Quantitative and temporal correlation between circulating cell-free Epstein-Barr virus DNA and tumor recurrence in nasopharyngeal carcinoma. *Cancer Res* 1999;59:5452-5.
- Le QT, Jones CD, Yau TK, Shirazi HA, Wong PH, Thomas EN, et al. A comparison study of different PCR assays in measuring circulating plasma Epstein-Barr virus DNA levels in patients with nasopharyngeal carcinoma. *Clin Cancer Res* 2005;11:5700-7.
- Lo YM, Chan AT, Chan LY, Leung SF, Lam CW, Huang DP, et al. Molecular prognostication of nasopharyngeal carcinoma by quantitative analysis of circulating Epstein-Barr virus DNA. *Cancer Res* 2000;60:6878-81.
- Lin JC, Wang WY, Chen KY, Wei YH, Liang WM, Jan JS, et al. Quantification of plasma Epstein-Barr virus DNA in patients with advanced nasopharyngeal carcinoma. *N Engl J Med* 2004;350:2461-70.
- Chan AT, Lo YM, Zee B, Chan LY, Ma BB, Leung SF, et al. Plasma Epstein-Barr virus DNA and residual disease after radiotherapy for

- undifferentiated nasopharyngeal carcinoma. *J Natl Cancer Inst* 2002;94:1614–9.
22. Stevens SJ, Verkuijlen SA, Hariwiyanto B, Harijadi, Fachiroh J, Paramita DK, et al. Diagnostic value of measuring Epstein-Barr virus (EBV) DNA load and carcinoma-specific viral mRNA in relation to anti-EBV immunoglobulin A (IgA) and IgG antibody levels in blood of nasopharyngeal carcinoma patients from Indonesia. *J Clin Microbiol* 2005;43:3066–73.
23. Wang WY, Twu CW, Chen HH, Jiang RS, Wu CT, Liang KL, et al. Long-term survival analysis of nasopharyngeal carcinoma by plasma Epstein-Barr virus DNA levels. *Cancer* 2013;119:963–70.
24. Shoukri MM, Asyali MH, Donner A. Sample size requirements for the design of reliability study: review and new results. *Stat Methods Med Res* 2004;13:251–71.
25. Lachin JM. The role of measurement reliability in clinical trials. *Clin Trials* 2004;1:553–66.

Clinical Cancer Research

An International Collaboration to Harmonize the Quantitative Plasma Epstein-Barr Virus DNA Assay for Future Biomarker-Guided Trials in Nasopharyngeal Carcinoma

Quynh-Thu Le, Qiang Zhang, Hongbin Cao, et al.

Clin Cancer Res 2013;19:2208-2215. Published OnlineFirst March 4, 2013.

Updated version Access the most recent version of this article at:
doi:[10.1158/1078-0432.CCR-12-3702](https://doi.org/10.1158/1078-0432.CCR-12-3702)

Supplementary Material Access the most recent supplemental material at:
<http://clincancerres.aacrjournals.org/content/suppl/2013/03/04/1078-0432.CCR-12-3702.DC1>

Cited articles This article cites 25 articles, 12 of which you can access for free at:
<http://clincancerres.aacrjournals.org/content/19/8/2208.full#ref-list-1>

Citing articles This article has been cited by 4 HighWire-hosted articles. Access the articles at:
<http://clincancerres.aacrjournals.org/content/19/8/2208.full#related-urls>

E-mail alerts [Sign up to receive free email-alerts](#) related to this article or journal.

Reprints and Subscriptions To order reprints of this article or to subscribe to the journal, contact the AACR Publications Department at pubs@aacr.org.

Permissions To request permission to re-use all or part of this article, use this link
<http://clincancerres.aacrjournals.org/content/19/8/2208>.
Click on "Request Permissions" which will take you to the Copyright Clearance Center's (CCC) Rightslink site.